Equation 19

$$\text{Slope} = \frac{\sum_{j=1}^{J}\left(\overline{R}_j - \overline{R}\right)\left(\overline{C}_j - \overline{C}\right)}{\sum_{j=1}^{J}\left(\overline{R}_j - \overline{R}\right)^2}$$

Equation 20

$$\text{Intercept} = \overline{C} - \text{slope} \times \overline{R}$$

(4) To pass this test, at each test site:

(i) The slope (calculated to at least 2 decimal places) must be in the interval specified for regression slope in table C–4 of this subpart; and

(ii) The intercept (calculated to at least 2 decimal places) must be in the interval specified for regression intercept in table C–4 of this subpart.

(iii) The slope and intercept limits are illustrated in figures C–2 and C–3 of this subpart.

(h) *Tests for comparison correlation.* (1) For each test site, calculate the (Pearson) correlation coefficient, r (not the coefficient of determination, r^2), using equation 21 of this section:

Equation 21

$$r = \frac{\sum_{j=1}^{J}\left(\overline{R}_j - \overline{R}\right)\left(\overline{C}_j - \overline{C}\right)}{\sqrt{\sum_{j=1}^{J}\left(\overline{R}_j - \overline{R}\right)^2 \sum_{j=1}^{J}\left(\overline{C}_j - \overline{C}\right)^2}}$$

(2) For each test site, calculate the concentration coefficient of variation, CCV, using equation 22 of this section:

Equation 22

$$\text{CCV} = \frac{1}{\overline{R}}\sqrt{\frac{\sum_{j=1}^{J}\left(\overline{R}_j - \overline{R}\right)^2}{J-1}}$$

(3) To pass the test, the correlation coefficient, r, for each test site must not be less than the values, for various values of CCV, specified for correlation in table C–4 of this subpart. These limits are illustrated in figure C–4 of this subpart.

[71 FR 61278, Oct. 17, 2006, as amended at 72 FR 32202, June 12, 2007]

TABLE C–1 TO SUBPART C OF PART 53—TEST CONCENTRATION RANGES, NUMBER OF MEASUREMENTS REQUIRED, AND MAXIMUM DISCREPANCY SPECIFICATIONS

Pollutant	Concentration range, parts per million (ppm)	Simultaneous measurements required				Maximum discrepancy specification, parts per mil-lion
		1-hour		24-hour		
		First set	Second set	First set	Second set	
Ozone	Low 0.06 to 0.10	5	6	....	....	0.02
	Med. 0.15 to 0.25	5	6	....	....	0.03
	High 0.35 to 0.46	4	6	....	....	0.04
	Total	14	18	....	....	
Carbon monoxide	Low 7 to 11	5	6	....	....	1.5
	Med. 20 to 30	5	6	....	....	2.0
	High 25 to 45	4	6	....	....	3.0
	Total	14	18	....	....	
Sulfur dioxide	Low 0.02 to 0.05	5	6	3	3	0.02
	Med. 0.10 to 0.15	5	6	2	3	0.03
	High 0.30 to 0.50	4	6	2	2	0.04
	Total	14	18	7	8	
Nitrogen dioxide	Low 0.02 to 0.08	....	....	3	3	0.02
	Med. 0.10 to 0.20	....	....	2	2	0.02
	High 0.25	....	....	2	2	0.03

Pollutant	Concentration range, parts per million (ppm)	Simultaneous measurements required				Maximum discrepancy specification, parts per million
		1-hour		24-hour		
		First set	Second set	First set	Second set	
Total		....	....	7	8	....

[75 FR 35601, June 22, 2010]

TABLE C–2 TO SUBPART C OF PART 53— SEQUENCE OF TEST MEASUREMENTS

Measurement	Concentration range	
	First set	Second set
1	Low	Medium.
2	High	High.
3	Medium	Low.
4	High	High.
5	Low	Medium.
6	Medium	Low.
7	Low	Medium.
8	Medium	Low.
9	High	High.
10	Medium	Low.
11	High	Medium.
12	Low	High.
13	Medium	Medium.
14	Low	High.
15	....	Low.
16	....	Medium.
17	....	Low.
18	....	High.

TABLE C–3 TO SUBPART C OF PART 53— TEST SPECIFICATIONS FOR PB IN TSP AND PB IN PM$_{10}$ METHODS

Concentration range equivalent to percentage of NAAQS in µg/m³.	30% to 250%
Minimum number of 24-hr measurements.	5
Maximum reference method analytical bias, D_q.	±5%
Maximum precision, P_R or P_C	≤15%
Maximum difference (D)	±20%
Estimated Method Detection Limit (MDL), µg/m³.	5% of NAAQS level.

[73 FR 67059, Nov. 12, 2008]

TABLE C–4 TO SUBPART C OF PART 53—TEST SPECIFICATIONS FOR PM$_{10}$, PM$_{2.5}$ AND PM$_{10–2.5}$ CANDIDATE EQUIVALENT METHODS

Specification	PM$_{10}$	PM$_{2.5}$			PM$_{10–2.5}$	
		Class I	Class II	Class III	Class II	Class III
Acceptable concentration range (R$_j$), µg/m³.	15–300	3–200 ..	3–200	3–200	3–200	3–200
Minimum number of test sites.	2	1	2	4	2	4
Minimum number of candidate method samplers or analyzers per site.	3	3	3[1]	3[1]	3[1]	3[1]
Number of reference method samplers per site.	3	3	3[1]	3[1]	3[1]	3[1]
Minimum number of acceptable sample sets per site for PM$_{10}$ methods:						
R$_j$ <60 µg/m³	3					
R$_j$ >60 µg/m³	3					
Total	10					
Minimum number of acceptable sample sets per site for PM$_{2.5}$ and PM$_{10–2.5}$ candidate equivalent methods:						
R$_j$ <30 µg/m³ for 24-hr or R$_j$ <20 µg/m³ for 48-hr samples.	....	3				
R$_j$ >30 µg/m³ for 24-hr or R$_j$ >20 µg/m³ for 48-hr samples.	3					
Each season	10	23	23	23	23	